

TITLE: <i>PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)</i>			
JCTLM WG2-P-03-B Version : 1.3	Date : Nov. 27, 2007	Authorized : JCTLM Executive	

***PROCESS FOR REVIEW OF
REFERENCE MEASUREMENT
SERVICES FROM LABORATORIES
THAT ARE ACCREDITED AS
CALIBRATION LABORATORIES
(ISO/IEC 17025 AND ISO 15195)***

**CRAIG M. JACKSON
ROBERT I. WIELGOSZ
LOTHAR SIEKMANN
LINDA THIENPONT**

**PROCEDURE
JCTLM WG2 P-03-B**

DRAFTING AUTHOR	AUTHORIZED
NAME	CRAIG M. JACKSON ROBERT I. WIELGOSZ LOTHAR SIEKMANN LINDA THIENPONT
FUNCTION	JCTLM WG2 PROCEDURES TEAM
SIGNATURE	

TITLE: <i>PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)</i>		
JCTLM WG2-P-03-B Version : 1.3	Date : Nov. 27, 2007	Authorized : JCTLM Executive

PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC-17025 + ISO-15195)

1. Contents

PARAGRAPH	TITLE	PAGE
1.	CONTENTS.....	2
2.	PURPOSE.....	2
3.	SCOPE.....	2
4.	ACRONYMS.....	3
5.	DEFINITIONS.....	3
6.	RESPONSIBILITIES AND AUTHORIZATIONS.....	3
7.	PROCEDURE.....	3
8.	RELATED DOCUMENTS.....	4
9.	REVISION HISTORY.....	5
10.	FLOWCHART	6

2. Purpose

This procedure describes the processes to be followed by review teams for reviewing nominations of reference measurement services provided by reference laboratories that are accredited according to ISO/IEC 17025 and ISO 15195 as calibration laboratories. Reference measurement services of laboratories not yet accredited will be reviewed according to procedure WG-2 P-03-C. Nominations from National Metrology Institutes with CMCs published in Appendix C of the BIPM KCDB according to CIPM MRA will have these reviewed according to WG-1 P-03-A.

3. Scope

This procedure will be applied by the review teams for assessment of laboratory services according to the criteria listed under ‘7. Procedure’ of this document based on the documentation provided by the laboratories and also regarding publicly available information on the laboratory performance (e.g. results of inter-comparison studies).

TITLE: <i>PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)</i>			
JCTLM WG2-P-03-B Version : 1.3	Date : Nov. 27, 2007	Authorized : JCTLM Executive	

4. Acronyms

Acronyms used in this procedure are defined in the document JCTLM WG-2 P-00, Outline on the Calibration and Measurement Hierarchy in Laboratory Medicine - Quality Policy and Definitions.

5. Definitions

Definitions for the terms used in this procedure are given in the document JCTLM WG-2 P-00, Outline on the Calibration and Measurement Hierarchy in Laboratory Medicine - Quality Policy and Definitions.

6. Responsibilities and Authorizations

Responsibilities and authorizations are given in the document JCTLM WG2-P-02

7. Procedure

- 7.1 The process for reviewing RML services by the Review Teams will take place annually. The JCTLM Secretariat sends received nominations to the JCTLM Review Teams after consultation with the WG2 chair. The appropriate review team will assess the nomination and other publicly available information according to the following criteria:
- 7.1.1 Review teams ascertain that the RML is accredited according to ISO/IEC 17025 **and** ISO 15195 as calibration laboratory on the basis of successful, and in general on-site, assessments by technical and management experts from accreditation bodies which are ILAC members (signatories). A copy of the laboratory's accreditation certificate shall be submitted to the JCTLM secretariat as part of the Reference Measurement Service nomination procedure.
 - 7.1.2 Review teams ascertain whether the Reference Method applied by the candidate laboratory is listed in the JCTLM Database. If the applied method has not yet been included in the JCTLM Reference Method list the laboratory will be notified to nominate the Reference Method for review and acceptance by JCTLM WG1.

If the Reference Method is not yet listed but submitted for listing at the same time, the JCTLM Executive will decide on the listing of the Method and the Reference Measurement Service at the same meeting.
 - 7.1.3 Review teams ascertain regular (at least annual) participation of the candidate laboratory in inter-laboratory comparisons provided by an appropriate proficiency testing/external quality assessment scheme for reference

TITLE: *PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)*

JCTLM WG2-P-03-B Version : 1.3	Date : Nov. 27, 2007	Authorized : JCTLM Executive	
-----------------------------------	-------------------------	---------------------------------	--

laboratories which fulfills the requirements of WG-2 P-02-8. The results of the inter-comparisons as well as the result of the individual laboratory related to its identity shall be publicly available.

It is also acceptable that a laboratory participates for another measurand of the same group of quantities (e.g., metabolites & substrates, low-molecular hormones, therapeutic drugs, enzymes) provided the same principle of measurement - e.g., IDMS, kinetic spectrophotometry for enzyme activities - is applied.

- 7.1.4 The results of the review are recorded in the submitted nomination Excel form, the review report form JCTLM WG2 P-03-F-02 and returned to the JCTLM Secretariat. The review report form is forwarded to the nominator at the end of the review cycle.
- 7.1.5 The Review Team recommendations and related documentation are forwarded to WG2 Chairs by the JCTLM Secretariat

8. Related documents

ISO 15195	In vitro diagnostic systems - Requirements for reference measurement laboratories
ISO/IEC 17025	General requirements for the competence of testing and calibration laboratories
JCTLM WG2-P-00	Outline of the Calibration and Measurement Hierarchy - Quality Policy and Definitions
JCTLM WG2-P-01	Outline of JCTLM Procedures for Identifying and Listing Reference Measurement Services of Reference Laboratories
JCTLM WG2-P-02	Process for Requesting and Accepting Nominations for JCTLM Listing of Reference Measurement Laboratory Services
JCTLM WG2-P03-A	Process for Review of Reference Measurement Services from National Metrology Institutes which are listed in Appendix C of the CIPM MRA for relevant CMCs.
JCTLM WG2-P-03-C	Process for Review of Reference Measurement Services from Laboratories which are Preparing for Accreditation
JCTLM WG2-P-03-F-02	Review Report form for a nominated Reference measurement Service
JCTLM WG2-P-04	Process for Listing Reference Measurement Laboratory Services
JCTLM WG2-P-05	Process for Delisting Reference Measurement Laboratory Services

TITLE: *PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)*

JCTLM WG2-P-03-B
Version : 1.3

Date :
Nov. 27, 2007

Authorized :
JCTLM Executive

9. Revision History

Version number	Date of Issue/Review	Summary of change
1.1	Dec. 06,2005	Document approved by Executive Meeting Nov 14, 2005
1.2	Dec. 01, 2006	Modification in 7.1.2
1.3	27 Nov. 2007	Modifications following decisions of the 5 th Executive Committee meeting

TITLE: PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES THAT ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO/IEC 17025 AND ISO 15195)

JCTLM WG2-P-03-B
Version : 1.3


Date :
Nov. 27, 2007

Authorized :
JCTLM Executive

10. Flowchart

PROCESS FOR REVIEW OF REFERENCE MEASUREMENT SERVICES FROM LABORATORIES WHO ARE ACCREDITED AS CALIBRATION LABORATORIES (ISO-17025/ 15195) WG-2 P-03-B

Responsibility: JCTLM Review Teams



Rev 05-12-06